

virion\serion	Analysten-zertifikat Hämolytisches System FB-IV-275	Version: A
Geltungsbereich	Institut Virion\Serion GmbH	Seite 1 / 1
Dokumentenname: FBIV275A_20240715_2426		

ANALYSENZERTIFIKAT / CERTIFICATE OF ANALYSIS

Produktname / Product name	Hämolytisches System für die KBR / Hemolytic System for the CFT	
Best.-Nr. / Order no.	HS-020, HS-023, HS-050, HS-053, HS-100, HS-103	
Chargen-Nr. / Lot no.	HS-020/2426, HS-023/2426, HS-050/2426, HS-053/2426, HS-100/2426, HS-103/2426	
Verfalldatum / Expiry Date	2024-08-25	
Volumen / Volume	20 mL, 3 x 20 mL, 50 mL, 3 x 50 mL, 100 mL, 3 x 100 mL	
Lagerung / Storage	2 – 8 °C	
Gebrauchsverdünnung / Working Dilution	gebrauchsfertig / ready to use	
Material / Material	Mischung aus 1 %iger Hammel-Erythrozyten-Suspension, Ambozeptor, KBR-Puffer und Stabilisatorlösung / Mixture of 1 % sheep erythrocytes, amboceptor, CFT buffer and stabilizing solution	
Aussehen / Appearance	Spezifikation / Specification	Flüssig, rot / liquid, red
	Resultat / Result	Flüssig, rot / liquid, red
Extinktion Überstand (at 570 nm) / Extinction supernatant (at 570 nm)	Spezifikation / Specification	≤ 0,040
	Resultat / Result	0,038
QC-Freigabedatum / QC-Release date	2024-07-15	

Unterschrift (Qualitätskontrolle)
Signature (quality control)